

Certificate of Analysis

Company: Doh Yey Smokieez
 1425 North 116 Rd
 Bristol, VT 05443

Sample ID: SLT0005-DY021123GUC-SCP

Lot: N/A

Report Date: 5/17/2023

Matrix: Flower

Date Analyzed: 5/16/2023

Customer ID: 220121-0

Date Sampled: N/A

Analyst: 011

Grower License #: SCLT0005
Date Received: 5/11/2023

Report ID: C230511A5

Cannabinoid Summary

Cannabinoid Profile	LOQ (mg/g)	Concentration (mg/g)	Weight (%)
CBDVA	0.0005	<LOQ	<LOQ
CBDV	0.0012	<LOQ	<LOQ
CBD	0.0008	0.73	0.07
CBGA	0.0008	5.00	0.50
CBG	0.0019	0.77	0.08
CBD	0.0019	<LOQ	<LOQ
THCV	0.0021	<LOQ	<LOQ
CBN	0.0013	<LOQ	<LOQ
Δ^9 -THC	0.0020	26.32	2.63
Δ^8 -THC	0.0019	<LOQ	<LOQ
THC-A	0.0034	193.55	19.35
CBC	0.0024	<LOQ	<LOQ
Total THC		196.06	19.61
Total CBD		0.64	0.06
Total Cannabinoids		226.37	22.64

19.61%
Total THC
0.06%
Total CBD
22.64%
**Total
Cannabinoids**
2.63%
 Δ^9 -THC
11.52%
**Percent
Moisture**
1 : 0
**THC : CBD
Ratio**

Cannabinoids Methodology: High Performance Liquid Chromatography (HPLC) using PerkinElmer FLEXAR™ with Photo Diode Array Detector (PDA)

Total CBD and total THC are calculated values, to account for assumed decarboxylation from the acid form (THCA or CBDA) to the neutral form, causing weight loss of the acid group. These values are calculated as follows:

Total THC = (THCA x 0.877) + Δ^9 -THC Total CBD = (CBDA x 0.877) + CBD

Ratio of Total CBD: Total THC Reagent Blanks: < LOQs for all analytes

LOQ = The lowest quantity that this method can reliably detect. Any cannabinoid that was not detected is assumed to be less than the stated LOQ (<LOQ).

All results reflect dry weight of material, based on % moisture of the sample.

Measurement of Uncertainty (MU): the parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the particular quantity subject to measurement.

Δ^9 -THC MU = $\pm 0.005\%$ Total THC MU = $\pm 0.007\%$

All other cannabinoid MU values are available upon request.

All moisture analysis is determined by loss-on-drying measurement using CHAUS Model M850 Moisture Content Readers.



This report shall not be reproduced except in full without approval of the laboratory. This is to provide assurance that parts of a report are not taken out of context. Results apply to the samples as received.

Certified by: *Luke E. M.*
 Luke Emerson Mason (Laboratory Director, Bia Diagnostics)